
# SHINWA EXPRESS COLL 38

## Analysis of doxylamine succinate using the ULTRON ES-OVM

Doxylamine succinate is used in the US mainly as a short-acting sleep aid. It is also prescribed in combination with vitamin B6 (pyridoxine hydrochloride) as a medication for relieving the symptoms of morning sickness during pregnancy. Shown below is a reversed-phase analysis of doxylamine succinate performed using the ULTRON ES-OVM.



 $\begin{array}{lll} \mbox{Column:} & \mbox{ULTRON ES-OVM (5 } \mbox{$\mu m$}) \\ \mbox{Column size:} & 150 \mbox{ mm L x } 4.6 \mbox{ mm I.D.} \\ \mbox{Mobile phase:} & 20 \mbox{ mM } \mbox{KH}_2\mbox{PO}_4 + 20 \mbox{ mM } \mbox{K}_2\mbox{HPO}_4 \mbox{ (pH } 6.0)/ \\ \end{array}$ 

 $CH_3CN = 90/10$ 1.0 mL/min  $\begin{array}{ll} \mbox{Column temp.:} & 25 \mbox{°C} \\ \mbox{Detection:} & UV-262 \mbox{ nm} \\ \mbox{Sample:} & 200 \mbox{ mg/L (in water)} \\ \end{array}$ 

Injection vol.: 2 μL

#### ● Features of the ULTRON ES-OVM

Flow rate:

- 1. Particularly effective for separations of chiral compounds.
- 2. Employs a highly denaturation resistant protein ligand\* (ovomucoid) as a stationary phase.
- 3. Displays a wide range of chiral recognition.
- 4. No sample pretreatment required for chiral separations. (Blood plasma samples can be analyzed without pretreatment.)
- 5. Can be used for reversed-phase analyses.
- 6. Trace analyses possible (several nanograms).
  - \*Ligand: A substance that binds selectively with a specific receptor.



50-2 Kagekatsu-cho, Fushimi-ku

Kyoto 612-8307 Japan

TEL: +81-75-621-2360 FAX: +81-75-602-2660

E-mail: info@shinwa-cpc.co.jp Website: http://shinwa-cpc.co.jp/en/

#### ■ Analytical column specifications

Product name	Particle size (µm)	Column dimensions Length × I.D. (mm)
ULTRON ES-OVM	5	150 × 2.0
ULTRON ES-OVM	5	150 × 4.6
ULTRON ES-OVM	5	150 × 6.0
ULTRON ES-OVM-C	5	150 × 4.6
ULTRON ES-OVM	10	250 × 4.6
ULTRON ES-OVM Prep (preparative)	10	250 × 20.0

### ■ Guard column specifications ■

Product name	Particle size (µm)	Column dimensions Length × I.D. (mm)
ULTRON ES-OVM.G	5	10 × 4.0
ULTRON ES-OVM Prep.G (preparative guard column)	10	15 × 8.0
ULTRON ES-OVM (guard cartridges (x2))	5	5 × 2.0
Guard cartridge holder (with adaptor)	For 5 × 2.0 mm columns	
ULTRON ES-OVM (guard cartridges (x 2))	5	10 × 4.6
Guard cartridge holder (with adaptor)	For 10 × 4.6 mm columns	

#### Protecting columns using guard columns

By using guard columns together with analytical columns, analytical columns can be protected from contamination, enabling longer column lifetimes. The use of guard columns is particularly recommended when analyzing samples containing large amounts of impurities.

Guard cartridges are also available when guard columns are changed frequently.

Please use guard columns with the same or smaller inner diameter than that of the analytical column being used.

#### ● ULTRON ES-OVM-C (new product)

Clopidogrel bisulfate evaluation conditions based on USP standards employed for shipping inspections.

(Performance reports included with columns.)

Please feel free to contact us with questions related to analyses. Column screening services are also available.

Please be aware that specifications and prices are subject to change without prior notification.